



Intertek
Conforms to UL STD 325
Control No. 3011624



brookfield industries, inc. NB-2000 Swinging Door Operator

Description:

The brookfield **NB-2000** Swinging Door Operator has been designed, tested and is manufactured by **brookfield industries, inc.** for use on linear accelerator swing doors weighing up to 20,000 lbs and 75” wide. The **NB-2000** can be expected to safely operate doors at the maximum weight for at least 500,000 cycles, providing none of the rated values are exceeded.

We are Authorized to Mark the **NB-2000** with the **ETL** and **CE** markings from Intertek Testing Services to assure compliance with UL 325, FCC Part 15 (Emissions) and the following International Standards: EN 60335-1 and -2 (Safety), EN 61000-6-2 (Immunity) and EN 61000-6-4 (Emissions). The “listing” reports specifically refer to the heavy commercial/industrial doors associated with linear accelerator rooms.

The **NB-2000** is an electro-mechanical power open/power close operator, which utilizes a state of the art PLC and positioning control components that allow us to offer special patent pending features such as *Absolute position feedback* and *Auto backlash removal* to name a few.

The PLC (Programmable Logic Control) is programmed by the manufacturer to accept input signals from push paddles, 2 and 3 button stations and partial open commands. Additional I/O (Input/Output) have been added to the **NB-2000** to accommodate any number of special customer requirements such as: E-stops, motion detectors, control mats, card readers, reversing edge sensors and battery backup to name a few.

Interface with the PLC is provided with a hand held, user-friendly operator panel to adjust door position presets (purchased separately).

Supply voltage for the **NB-2000** is normally 115 VAC 50/60 Hz. For European service, we offer a 230 VAC 50/60 Hz hookup complete with a NEMA 1 enclosure and 0.5KVA 230VAC to 120 VAC step down transformer.

Rating a Door Operator:

The rating of any door operator in a particular application cannot be based solely on the weight and width of the door. Other factors such as an “out-of plum” frame, “hinge bind”, hinge friction, acceleration and deceleration rates or pressure differential from one side of the door to the other, may have a substantial affect on the total operating torque it takes to move the door at ANSI speeds. For Example, a 4500 lb, 60” wide door would be well within the weight and width limitations; however, if any of the other factors are not properly controlled, the total operating torque could exceed the rated values. In the design and testing of the NB-2000 door operator, we have factored these variables into the Rated Operating Torque. This assures the customer they are getting the most dependable product at a reasonable cost.

Maximum Rated Operating Torque =	3,000 lb-in (339 N-m)
Maximum Rated Door Weight =	20,000 lbs (9080 kg)
Maximum Rated Door Width =	75 in (1.91m)

The NB-2000 can be expected to operate any linear accelerator swing door within the rated values for at least 500,000 cycles, provided the door is operating within the ANSI 156.10 speed criteria and the acceleration rate for a particular door weight has not been exceeded.



Specification:

- 1) **Forward and reverse torque adjustment:** via current limit trimpot settings on the motor control to adjustment the operating torque in order to comply with the entrapment protection requirements of UL 325. This eliminates the need for unreliable and hard to adjust manual clutches or electromagnetic clutches that are not designed for slippage or stall applications.

- 2) **Automatic egress:** or door reversal whenever entrapment protection has been violated in either the opening or closing directions; however, door shall not reverse automatically when in the latchcheck or “creep close” mode. This feature can be adjustable for time delay and reversal force.
- 3) **Auto backlash removal:** automatically closes the door and removes all mechanical backlash in the system when in the latchcheck “creep close” mode. If a pressure differential exists or if the door wants to float open slightly, the motor will restart and “clamp” the door against the frame; thus, assuring the frame interlocks will not be disengaged and the response time to open will be optimized.
- 4) **Absolute position feedback control:** this assures the CPU always knows the door’s position. During installation, a power interruption, or if electrical noise is encountered, the door is not required to be “homed”, “reset” nor go through a “learn speed cycle” at any time.
- 5) **Manual operation:** shall be in accordance with UL 325 during a power loss.
- 6) **Supply voltage:** 115VAC +/- 10% (230VAC for European service) 50/60 Hertz single phase. In-Line circuit breakers supplied with motor control and PLC. Surge protection, line filters, and EMI ferrites shall be included.
Current Consumption: maximum 5 amperes
- 7) **PLC/Logic Control:**
 - a) Shall be a PLC with sufficient I/O and a CPU (Central Processing Unit) with adequate memory, response times and scanning rates in order to properly control the motion and positioning of linear accelerator swing Doors.
 - b) Outputs commands shall be the internal type, integral with the PLC. No external limit switches shall be allowed for control of door positioning.
 - c) A means to interface with the PLC for adjusting preset values for the open, partial open, closed, latchcheck and backcheck positions.
 - d) Diagnostics and troubleshooting of the PLC shall be provided with LED and modular plug-in components.
 - e) The PLC shall be provided with an internal battery to store the door position presets in the CPU memory.
- 8) **Motor:** 1/8 hp permanent magnet 90 volt DC motor 1750 RPM TENV
- 9) **Motor Control:** shall be a full-wave, four quadrant, regenerative, 90 VDC variable speed control with the following functions:

FWD/REV maximum speed	FWD/REV current limit	IR compensation
FWD/REV acceleration/deceleration	1% speed regulation	50:1 speed range.
- 10) **Speed Control:** a means of controlling independent forward and reverse speeds per ANSI 156.10 as well as controlling latchcheck and backcheck “creep” speeds. This can be accomplished externally with speed pots or internally with the PLC.
- 11) **Drive train and linkage mechanism:** shall be designed to allow manual operation of the door per UL 325 in addition to assuring each component from the motor to the door attachment point is properly “sized” in order to transfer all operating torques and forces as defined for Linear Accelerator Swing Doors. Standard linkage shall consist of a pull open design with crank arm, slider block and cam follower assembly.
- 12) **Enclosure:** U- shaped 16 gage cover with 1/8” thick end plates reinforced with angle iron. Overall 55” x 11-1/8” x 11-1/8”. (2) 7/8” dia. penetrations for 1/2” conduit on each side or metric equivalent for European installations. For 230VAC service, the transformer shall be housed in a separate 16” x 20” x 6-5/8” NEMA 1 vented enclosure. Penetrations are provided for 1/2” conduit or the metric equivalent for European installations.
- 13) **Materials:** ASTM A569, A36, AISI 1018 cold rolled steel, grade 5 bolting or better.
- 14) **Mounting hardware:** the NB-2000 shall be mounted with (6) 3/8” grade 5 diameter bolts with compatible washers and lock washers. Hardware must also be properly tightened with adequate thread engagement.
- 15) **Finish:** all exposed metal surfaces shall be prime painted.
- 16) **Functionality test:** each NB-2000 is cycle tested in position for 24 hrs. prior to shipment. Each unit is checked for leaks and that all I/O are functioning properly.
- 17) **Installation:** the NB-2000 shall be installed per drawing NB-2000-22A for pull open linkages and O31299K2 and O30299K2 for push open applications.
- 18) **Battery Backup (optional):** Opens the door during power interruption only. A 12VDC, 1.2 Ah battery with float charger and test switch shall be assembled in a grounded and vented 6” x 12” x 14” NEMA 1 enclosure. (3) holes are provided at the top of the enclosure each for 1/2” conduit or the metric equivalent for European installations. The first shall be used for a 110V AC line from the power source, the 2nd for a 110VAC hookup to the door operator and the 3rd for a 12VDC hookup to the door operator. An end of travel limit switch shall also be provided.

Rev.	Revision note	Date	Sign	Check



Notes:

Tolerances	
Decimal	+/- 0.015
Fractional	+/- .xxx
Angular	+/- .xxx


- NOTES:**
- LINKAGE ARM ADJUSTMENT +/- 1/4" SHIM PACK
THIS ALLOWS FOR PROPER ENGAGEMENT OF CAM FOLLOWER INTO SLIDER BLOCK AND PREVENTS THE LINKAGE ARM FROM INTERFERING WITH THE DOOR OR THE SLIDER BLOCK.
 - SLIDER BLOCK AND CAM FOLLOWER ASSEMBLY. 13/32" DIA. HOLES PROVIDED FOR 3/8" SOCKET HEAD BOLTS. (TYP. 2 PLACES)
 - ALL DIMENSIONS IN INCHES.
 - 1/2" DIA. THRU HOLES FOR 3/8" DIA. GRADE 5 BOLTS. (TYP. 6 PLACES)
 - FIELD WIRING (LH OR RH): ALLWAYS RUN 110VAC THRU LEFT SIDE OF COVER AND CLASS 2 (24VDC OR LESS) THRU RIGHT SIDE OF COVER.
(2) 7/8" DIA. HOLES FOR 1/2" DIA. CONDUIT (TYP. BOTH SIDES OF COVER)

	NB-2000-RH MAIN INSTALLATION		Drawn by KPK
	Material		Checked by
	Date 12-16-13	Scale 1:1	
	PULL OPEN LINKAGE brookfield industries inc. 99 West Hillside Ave. : Thomaston, CT P: (860) 283-6211 F: (860) 283-6123 info1@brookfieldindustries.com		

Rev.	Revision note	Date	Sign	Check

A
B
C
D
E
F

A
B
C
D
E
F




$L = \sqrt{J^2 + 68.82J + 1184}$
 $L =$ _____

Notes:	
NOTES:	
1) 15.0 MAX DIMENSION. IF EXCEEDED, CONTACT ENGINEERING DEPT.	
2) FOR DOORS LESS THAN 48" WIDE, THE QUICK DISCONNECT LINKAGE OPTION MAY BE REQUIRED TO MEET THE 50 LB FORCE FOR MANUAL OPERATION WITH POWER REMOVED.	
3) OUTPUT SHAFT ROTATION=96 DEG. FOR 90 DEG DOOR SWING	
Tolerances	
Decimal	
Fractional	+/- .xxx
Angular	+/- .xxx

	NB-2000 MAIN-SUB INSTALLATION		Drawn by KPK
			Checked by _____
	Material _____	Date 10-18-13	Scale 1:1
	PUSH OPEN LINKAGE FOR DOUBLE DOOR LEAD/FOLLOW SEQUENCING brookfield industries inc. 99 West Hillside Ave. : Thomaston, CT P: (860) 283-6211 F: (860) 283-6123 info1@brookfieldindustries.com		


1 2 3 4 5 6 7 8

Rev.	Revision note	Date	Sign	Check




NOTES:

- 1) LINKAGE ARM PROVIDED WITH SHIM PACK WITH +/- 1/4" ADJUSTMENT
- 2) 13/32" DIA. HOLES PROVIDED FOR 3/8" DIA. BOLTS (TYP.2), (Grade 5).
- 3) ALL DIMENSIONS IN INCHES.
- 4) 1/2" DIA. THRU HOLES FOR 3/8" DIA. GRADE 5 BOLTS(TYP. 6 PLACES)
- 5) FIELD WIRING (LH OR RH): ALWAYS RUN 110VAC THRU LEFT SIDE OF COVER AND CLASS 2 (24VDC OR LESS) THRU RIGHT SIDE OF COVER.
(2) 7/8" DIA. HOLES FOR 1/2" DIA. CONDUIT (TYP. BOTH SIDES OF COVER)



	NB-2000 (SUB) LHR OPERATOR		Drawn by <i>kpk</i>
	Material		Checked by
	Date 12-17-13	Scale 1:1	
	INSTALLATION DRAWING FOR PUSH OPEN LINKAGE		
brookfield industries inc.		www.brookfieldindustries.com	
99 West Hillside Ave. : Thomaston, CT			
P: (860) 283-6211 F: (860) 283-6123			
info1@brookfieldindustries.com		Drawing No./File Name	
		NB-2000-MS6	

Tolerances	Decimal	Fractional	Angular	Rev.	Revision note	Date	218	Sign	Check
				1	DO Item numbers added	7/04			




SPEED BOARD LEGEND

PRESET 1 CREEP CLOSE (REV) SPEED POT

PRESET 2 CREEP OPEN (FWD) SPEED POT

PRESET 3 CLOSE (REV) SPEED POT

PRESET 4 OPEN (FWD) SPEED POT



NEMA MOUNTING PLATE

	Components & Mounting Plate		Drawn by kpk
	Material		Checked by
	Date 5/20/03		Scale 1:1
	NB-500, -1000, -2000 Logic Assembly brookfield industries inc. 99 West Hillside Ave.: Thomaston, CT P: (860) 283-6211 F: (860) 283-6123 info@brookfieldindustries.com		

2 3 4 5 6 7 8

RevNo	Revision note	Date	Signature	Checked

A

B

C

D

E

F

A

B


C

D


E

F

TOP VIEW



FRONT VIEW



SIDE VIEW

DO 605
LIMIT SWITCH

Drawing Notes:

INSTALL ONLY
WITH BATTERY
BACKUP SYSTEMS.

Tolerances

Decimal +/- 0.015 UNO

Fractional +/- .xxx

Angular +/- .xxx

NB-500, NB-1000, NB-2000

Drawn by KK
Checked by XXX

LIMIT SWITCH ASSY/ BATTERY BACKUP

Edition 0

Approved by - date
XXX - 00/00/00 XXX

Filename

Date
1/26/06

Scale
1:1

brookfield industries inc.
99 West Hillside Ave. : Thomaston, CT 06787
P: (860) 283-6211 F: (860) 283-6123

Material

DO2506KI

6 7 8